

Easybeat Display Messages

PROB OFF	probe is not on a finger
PROB DIS	probe is disconnected
PULS LST	pulse has been lost
PROB FND	pulse has been found, and reading will show soon
MOVE FND	movement detected, please ensure finger probe remains as stationary as possible